



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/597,829

08/09/2006

Russell Keene

W-354-02

6779

43840 7590 12/16/2008  
WATERS INVESTMENTS LIMITED  
C/O WATERS CORPORATION  
34 MAPLE STREET - LG  
MILFORD, MA 01757

EXAMINER

DUNWOODY, AARON M

ART UNIT

PAPER NUMBER

3679

MAIL DATE

DELIVERY MODE

12/16/2008

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/597,829	<b>Applicant(s)</b> KEENE, RUSSELL	
	<b>Examiner</b> Aaron M. Dunwoody	<b>Art Unit</b> 3679	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 09 August 2008.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-10 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-10 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____  | 6) <input type="checkbox"/> Other: _____                          |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-10 are rejected under 35 U.S.C. 103(a) as being unpatentable over US patent 6102449, Welsh in view of US patent 5848813, Albrecht.

In regards to claims 1-10, Welsh discloses a coupling element, comprising:

a male sealing element having a first end, second end, and a longitudinal axis extending between said first end and said second end, wherein said male sealing element has a generally cylindrical shape, wherein said male sealing element defines a fluid passageway therethrough for the transmission of fluid, wherein said male sealing element is slideably coupled to a ferrule (160), wherein said first end defines a conical sealing surface, wherein a female sealing element (132) defines a complementary conical geometry; and

a biasing element (154) disposed between a retaining ring (170) and said ferrule for biasing said first end into direct abutting contact with said female sealing element with a biasing force sufficient to form a fluid-tight seal between said first end and said female sealing element.

Welsh does not disclose the conical sealing surface having a mismatched angle to the female sealing element. Albrecht teaches a conical sealing surface (32) having a

Art Unit: 3679

mismatched angle to the female sealing element (22) "to ensure that said seal surface of the male portion seats on an inner diameter of the seal surface of the female portion so that fluid sealing takes place at a minimum diameter of said mating seal surfaces" (col. 2, lines 20-24). Since both Welsh and Albrecht are concerned with tapered sealing surfaces, it would have been obvious to one having ordinary skill in the art at the time the invention was made to fabricate the conical sealing surface with a mismatched angle to the female sealing element to ensure that said seal surface of the male portion seats on an inner diameter of the seal surface of the female portion so that fluid sealing takes place at a minimum diameter of said mating seal surfaces, as taught by Albrecht.

### ***Response to Arguments***

Applicant's arguments filed 9/8/2008 have been fully considered but they are not persuasive.

In response to applicant's argument that obviousness of Albrecht is misleading, the test for obviousness is not whether the features of a secondary reference may be bodily incorporated into the structure of the primary reference; nor is it that the claimed invention must be expressly suggested in any one or all of the references. Rather, the test is what the combined teachings of the references would have suggested to those of ordinary skill in the art. See *In re Keller*, 642 F.2d 413, 208 USPQ 871 (CCPA 1981).

In response to applicant's argument that the examiner's conclusion of obviousness is based upon improper hindsight reasoning, it must be recognized that any judgment on obviousness is in a sense necessarily a reconstruction based upon hindsight reasoning. But so long as it takes into account only knowledge which was

Art Unit: 3679

within the level of ordinary skill at the time the claimed invention was made, and does not include knowledge gleaned only from the applicant's disclosure, such a reconstruction is proper. See *In re McLaughlin*, 443 F.2d 1392, 170 USPQ 209 (CCPA 1971).

In response to applicant's argument that there is no suggestion to combine the references, the examiner recognizes that obviousness can only be established by combining or modifying the teachings of the prior art to produce the claimed invention where there is some teaching, suggestion, or motivation to do so found either in the references themselves or in the knowledge generally available to one of ordinary skill in the art. See *In re Fine*, 837 F.2d 1071, 5 USPQ2d 1596 (Fed. Cir. 1988) and *In re Jones*, 958 F.2d 347, 21 USPQ2d 1941 (Fed. Cir. 1992). In this case, since both Welsh and Albrecht are concerned with tapered sealing surfaces, it would have been obvious to one having ordinary skill in the art at the time the invention was made to fabricate the conical sealing surface with a mismatched angle to the female sealing element to ensure that said seal surface of the male portion seats on an inner diameter of the seal surface of the female portion so that fluid sealing takes place at a minimum diameter of said mating seal surfaces, as taught by Albrecht.

In response to applicant's argument that Albrecht is nonanalogous art, it has been held that a prior art reference must either be in the field of applicant's endeavor or, if not, then be reasonably pertinent to the particular problem with which the applicant was concerned, in order to be relied upon as a basis for rejection of the claimed invention. See *In re Oetiker*, 977 F.2d 1443, 24 USPQ2d 1443 (Fed. Cir. 1992). In this case,

Art Unit: 3679

Albrecht is pertinent to the particular problem of tapered sealing surfaces with which the applicant is concerned.

In response to applicant's arguments against the references individually, one cannot show nonobviousness by attacking references individually where the rejections are based on combinations of references. See *In re Keller*, 642 F.2d 413, 208 USPQ 871 (CCPA 1981); *In re Merck & Co.*, 800 F.2d 1091, 231 USPQ 375 (Fed. Cir. 1986).

Applicant argues that Welsh fails to disclose a ferrule. The Examiner disagrees. A ferrule is a usually metal sleeve used especially for joining or binding one part to another (as pipe sections or the bristles and handle of a brush) [Merriam-Webster Online Dictionary copyright © 2008 by Merriam-Webster, Incorporated]. Clearly, Welsh discloses a ferrule defined as a sleeve used especially for joining or binding one part to another.

In response to applicant's argument that the references fail to show certain features of applicant's invention, it is noted that the features upon which applicant relies (i.e., a male sealing element with a first end for sealing engagement to a female sealing element at a terminal end) are not recited in the rejected claim(s). Although the claims are interpreted in light of the specification, limitations from the specification are not read into the claims. See *In re Van Geuns*, 988 F.2d 1181, 26 USPQ2d 1057 (Fed. Cir. 1993).

Further, it appears Applicant is confusing "end" with "terminal end".

In response to Applicant's argument that Welch includes additional structure not required by Applicant's invention, it must be noted that Welch in view of Albrecht

Art Unit: 3679

disclose the invention as claimed. The fact that Welch discloses additional structure not claimed is irrelevant.

Further, a recitation with respect to the manner in which an apparatus is intended to be employed does not impose any structural limitation upon the claimed apparatus which differentiates it from a prior art reference disclosing the structural limitations of the claim. In re Pearson, 494 F.2d 1399, 181 USPQ 641 (CCPA 1974); In re Yanush, 477 F.2d 958, 177 USPQ 705 (CCPA 1973); In re Finsterwalder, 436 F.2d 1028, 168 USPQ 530 (CCPA 1971); In re Casey, 370 F.2d 576, 152 USPQ 235 (CCPA 1967); In re Otto, 312 F.2d 937, 136 USPQ 458 (CCPA 1963); Ex parte Masham, 2 USPQ2d 1647 (BdPatApp & Inter 1987).

### ***Conclusion***

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Art Unit: 3679

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Aaron M. Dunwoody whose telephone number is 571-272-7080. The examiner can normally be reached on 7:30 am - 4:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Daniel P. Stodola can be reached on 571-272-7087. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Aaron M Dunwoody/  
Primary Examiner, Art Unit 3679

.amd